





Agilent Technologies

Certificate of Conformance



Agilent Technologies, Inc.
Dissolution Systems
13000 Weston Parkway
Cary, NC 27513
Phone: +1 919.677.1108
Fax: +1 919.677.1138

12-5149V - Verified, Vessel, 1L TruAlign W/Collar
Part Number - Part Description

100276
Serial Number

Clear Glass
Material

The accuracy of these parts was determined by reference comparison to working standards traceable to the National Institute of Standards and Technology (NIST). Any limitations or remarks pertaining to this instrument or calibration are shown below.



Ref No.	USP Spec	Inspection Range	Actual
1	98-106	98-106	101.359
2	160-210	160-210	181.372
3	49-53	49-53	49.849

All readings are in mm

Procedure Used:	G. Inspection
Temperature/RH:	19.9oC/35.0%
Uncertainty:	(3.3+0.1L)µm

Equipment Used:	Instrument	Model Number	Serial Number	Accuracy
	Zeiss CMM	UMC 850	74087	+/- 5µm

R. BOUJLIL
Inspected By Vendor

[Signature]
Signature

02/17/2011
Date

David Stearns
Certified By

[Signature]
Signature

02/28/11
Date

This certificate shall not be reproduced without written approval of Agilent Technologies, Inc.
K4350-00025A COC.1L Dissolution Vessel.USB.docx

