









NAUTILUS
DELTA™ | TIP
CONFIRMATION
SYSTEM

INSTRUCTIONS FOR USE

LIST OF ALL APPLICABLE ICONS

	Manufacturer
	Consult instructions for use
Rx Only	Caution: Federal (U.S.A.) law restricts this device to sale by or on the order of a physician
	Serial Number
	Date of Manufacture
	Direct Current
	Do not dispose of battery with ordinary municipal waste








	Type CF applied part
	Do not dispose with ordinary municipal waste
	The Green Dot
	Catalog Number
	Caution
	Class II Electrical Equipment
	Operate Indoors Only

Table of Contents

1.0 NAUTILUS DELTA™ Tip Confirmation System (TCS) Description	4
2.0 Indications for Use	6
3.0 Contraindications	6
4.0 Warnings	6
5.0 Precautions	9
6.0 Assembling the NAUTILUS DELTA™ TCS	
6.1 Attaching the Patient Module to the Roll Stand	10
6.2 Connecting the Patient Module to the Display	10
6.3 Connecting the ECG Cable to the Patient Module	10
7.0 NAUTILUS DELTA™ TCS Information	
7.1 NAUTILUS DELTA™ TCS Graphical Interface, Controls and Indicators	11
7.2 Shutdown Menu	16
7.3 Settings Icon	16
7.3.1 Settings – General	16
7.3.2 Settings – Printers	17
7.3.3 Settings – File Management	18
8.0 NAUTILUS DELTA™ TCS Catheter Guidance	19
9.0 Troubleshooting and Error Messages	
9.1 Errors Screens	21
9.2 ECG Trouble Shooting	22
10.0 Battery Installation and Removal	
10.1 Installation	25
10.2 Removal	25
11.0 Remote Control Installation	26
12.0 NAUTILUS DELTA™ TCS Printer Installation	26
13.0 Cleaning and Disinfection	
13.1 Cleaning	26
13.2 Disinfection	27
14.0 Warranty	27
15.0 Service and Repair	28
16.0 Technical Specifications	28
17.0 Disposal Information	28
18.0 Deleting Patient Information	28
19.0 Manufacturer's Declaration	29